

Air Ionizer Verification Record

Ionizer Verification Sequence Number: 08-094

WORKING STANDARD USED						
Asset/ISO #:	Manufacturer:	Model:	Serial No.	Calibration Date:	Calibration Due:	Calibration By:
23845	ION	775	6778	8-30-07	8-30-08	JPL

AIR IONIZER INFORMATION						
Asset/ISO #:	Manufacturer:	Model:	Serial No.	Verification Date:	Verification Due:	Verification By:
28300	SIMCO	Aerostat R	33674327T	7-31-08	1-11-09	(38)
Inspector:	Location:	Owner:	Fail: Y/N ?	Cleaned: Y/N ?	Adjusted: Y/N ?	Prior Sequence#
Minh Do	317/210	John M.	N	N	N	08-090

VERIFICATION DATA						
HBM Sensitivity Level: <u>50 V</u> (from Table 1)						
Fan controller setting: <u>High</u> (High, Low, NA)						
Distance of ionizer from the charge plate: <u>30"</u>						
Ionizer Float Potential Tolerance \pm <u>50</u> Vdc. (from Table 1)						
Measured Float Potential values recorded below.						
1	2	3	4	5	Comments:	
<u>0</u> Vdc.	<u>0</u> Vdc.	<u>0</u> Vdc.	<u>0</u> Vdc.	Vdc.		
Ionizer Discharge Voltage Range: \pm 1000 Vdc to $< \pm$ <u>50</u> Vdc (from Table 1)						
Ionizer Discharge Time Tolerance: <u>20</u> seconds. (from Table 1)						
Measured Discharge Time in second(s) and recorded values below.						
1 (+1000 to +Vdc)	2 (+1000 to +Vdc)	3 (+1000 to +Vdc)	4 (+1000 to +Vdc)	5 (+1000 to +Vdc)	Comments:	
<u>4.9</u> sec	<u>5.2</u> sec	<u>4.5</u> sec	<u>5.0</u> sec	<u>5.2</u> sec		
1 (-1000 to -Vdc)	2 (-1000 to -Vdc)	3 (-1000 to -Vdc)	4 (-1000 to -Vdc)	5 (-1000 to -Vdc)	Comments:	
<u>6.8</u> sec	<u>6.4</u> sec	<u>6.5</u> sec	<u>6.6</u> sec	<u>7.2</u> sec		

Record any corrective action required to restored ionizer operation (cleaning, adjustment, replacement, etc.)

If Ionizer was replaced, indicate below the identification of replacement.

Asset/ISO #: _____ Manufacturer: _____ Model: _____ Serial No.: _____

Sequence number for verification of replacement ionizer: _____

Record inspection schedule and rational for that schedule.